

835MHz Brain Dipole Validation

SAM Phantom; FlatSection; Probe: ET3DV6 - SN1617 -- ConvF(6.40,6.40,6.40)


Head 835 MHz: $\sigma = 0.89$ mho/m $\epsilon_r = 41.2$ $\rho = 1.00$ g/cm³; Crest factor: 1.0

SAR (1g): 2.52 mW/g

835MHz Brain Dipole Validation (D835V2 S/N:451)

Frequency: 835 MHz; Antenna Input Power: 250 [mW]; Ambient Temp. (°C) - 21.6; Tissue Temp. (°C) - 20.7

Brain Tissue Simulating Liquid [06/19/2002]



835MHz Brain Dipole Validation

SAM Phantom; FlatSection; Probe: ET3DV6 - SN1617 -- ConvF(6.40,6.40,6.40)

Head 835 MHz: $\sigma = 0.89$ mho/m $\epsilon_r = 40.3$ $\rho = 1.00$ g/cm³; Crest factor: 1.0

SAR (1g): 2.56 mW/g

835MHz Brain Dipole Validation (D835V2 S/N:451)

Frequency: 835 MHz; Antenna Input Power: 250 [mW]; Ambient Temp. (°C) - 20.7; Tissue Temp. (°C) - 20.0

Brain Tissue Simulating Liquid [06/20/2002]

