

SAR Data Report 02041109

Start : 20-May-02 10:09:19 am
End : 20-May-02 10:20:57 am
Code Version : 4.08
Robot Version: 4.08

Product Data:

Type : Validation
Frequency : 835 MHz
Transmit Pwr : 0.250 W
Antenna Type : Dipole

Measurement Data:

Phantom Name : SAM-FLAT
Phantom Type : Uniphantom
Tissue Type : Brain
Tissue Dielectric : 40.600
Tissue Conductivity : 0.910
Tissue Density : 1.000
Robot Name : CRS

Probe Data:

Probe Name : PCT002
Probe Type : E Fld Triangle
Frequency : 835 MHz
Tissue Type : Brain
Calibrated Dielectric : 40.700
Calibrated Conductivity : 0.890
Calibrated Density : 1.000
Probe Offset : 2.400 mm
Conversion Factor : 5.800
Probe Sensitivity : 3.597 3.474 3.049 mV/(mW/cm²)
Amplifier Gains : 20.00 20.00 20.00

Sample:

Rate: 6000 Samples/Sec
Count: 100 Samples
NIDAQ Gain: 5

Comments:

835 MHz Validation

CF=1; Amb. Temp= 21.3 'C; Liq. Temp=21.0 'C

Area Scan - Max Peak SAR Value at x=4.0 y=-1.0 = 2.34 W/kg

Zoom Scan - Max Peak SAR Value at x=4.0 y=-4.0 z=0.0 = 3.65 W/kg


Max 1g SAR at x=4.0 y=-2.0 z=0.0 = 2.44 W/kg

Max 10g SAR at x=4.0 y=-1.0 z=0.0 = 1.53 W/kg


Validation Results at 0.25 W:


Peak Nominal = 3.5, Error: 4.37 %
1g Nominal = 2.4, Error: 2.94 %
10g Nominal = 1.6, Error: -1.25 %

SAR - Z Axis
at Hotspot x:4.0 y:-4.0



1g SAR Values





SAR Data Report 02031519

Start : 20-May-02 09:51:38 am
End : 20-May-02 10:02:17 am
Code Version : 4.08
Robot Version: 4.08

Product Data:

Type : Validation
Frequency : 1900 MHz
Transmit Pwr : 0.250 W
Antenna Type : Dipole

Measurement Data:

Phantom Name : SAM-FLAT
Phantom Type : Uniphantom
Tissue Type : Brain
Tissue Dielectric : 40.200
Tissue Conductivity : 1.390
Tissue Density : 1.000
Robot Name : CRS

Probe Data:

Probe Name : PCT002
Probe Type : E Fld Triangle
Frequency : 1900 MHz
Tissue Type : Brain
Calibrated Dielectric : 40.200
Calibrated Conductivity : 1.410
Calibrated Density : 1.000
Probe Offset : 2.400 mm
Conversion Factor : 4.700
Probe Sensitivity : 3.000 2.995 2.653 mV/(mW/cm^2)
Amplifier Gains : 20.00 20.00 20.00

Sample:

Rate: 6000 Samples/Sec
Count: 100 Samples
NIDAQ Gain: 5

Comments:

1900 MHz Validation

Amb. Temp= 22.5 'C; Liq. Temp=21.4 'C

Area Scan - Max Peak SAR Value at x=1.0 y=-2.0 = 9.58 W/kg

Zoom Scan - Max Peak SAR Value at x=1.0 y=-2.0 z=0.0 = 17.37 W/kg


Max 1g SAR at x=1.0 y=-1.0 z=0.0 = 10.29 W/kg

Max 10g SAR at x=1.0 y=-1.0 z=0.0 = 5.39 W/kg

Validation Results at 0.25 W:

Peak Nominal = 18.0, Error: -3.64 %
1g Nominal = 9.9, Error: 3.67 %
10g Nominal = 5.1, Error: 5.15 %

SAR - Z Axis
at Hotspot x:1.0 y:-2.0



1g SAR Values

