

SAR Data Report 02071801

Start : 18-Jul-02 08:49:41 am
End : 18-Jul-02 08:55:39 am
Code Version : 4.08
Robot Version: 4.08

Product Data:

Type : Validation
Frequency : 835 MHz
Transmit Pwr : 0.250 W
Antenna Type : Dipole

Measurement Data:

Phantom Name : SAM-FLAT
Phantom Type : Uniphantom
Tissue Type : Brain
Tissue Dielectric : 40.560
Tissue Conductivity : 0.910
Tissue Density : 1.000
Robot Name : CRS

Probe Data:

Probe Name : PCT002
Probe Type : E Fld Triangle
Frequency : 835 MHz
Tissue Type : Brain
Calibrated Dielectric : 40.700
Calibrated Conductivity : 0.890
Calibrated Density : 1.000
Probe Offset : 2.400 mm
Conversion Factor : 5.800
Probe Sensitivity : 3.597 3.474 3.049 mV/(mW/cm^2)
Amplifier Gains : 20.00 20.00 20.00

Sample:

Rate: 6000 Samples/Sec
Count: 100 Samples
NIDAQ Gain: 5

Comments:

835 MHz Validation

CF=1; Amb. Temp= 22.4 'C; Liq. Temp=22.1 'C

Area Scan - Max Peak SAR Value at x=0.0 y=-3.0 = 2.24 W/kg

Zoom Scan - Max Peak SAR Value at x=1.0 y=-1.0 z=0.0 = 3.47 W/kg


Max 1g SAR at x=1.0 y=-3.0 z=0.0 = 2.37 W/kg

Max 10g SAR at x=1.0 y=-3.0 z=0.0 = 1.49 W/kg


Validation Results at 0.25 W:


Peak Nominal = 3.5, Error: -1.67 %
1g Nominal = 2.4, Error: -0.38 %
10g Nominal = 1.6, Error: -3.72 %

SAR - Z Axis
at Hotspot x:1.0 y:-1.0



1g SAR Values





SAR Data Report 02071901

Start : 19-Jul-02 8:20:30 am
End : 19-Jul-02 8:26:10 am
Code Version : 4.08
Robot Version: 4.08

Product Data:

Type : Validation
Frequency : 1900 MHz
Transmit Pwr : 0.250 W
Antenna Type : Dipole

Measurement Data:

Phantom Name : SAM-FLAT
Phantom Type : Uniphantom
Tissue Type : Brain
Tissue Dielectric : 41.110
Tissue Conductivity : 1.420
Tissue Density : 1.000
Robot Name : CRS

Probe Data:

Probe Name : PCT002
Probe Type : E Fld Triangle
Frequency : 1900 MHz
Tissue Type : Brain
Calibrated Dielectric : 40.200
Calibrated Conductivity : 1.410
Calibrated Density : 1.000
Probe Offset : 2.400 mm
Conversion Factor : 4.700
Probe Sensitivity : 3.000 2.995 2.653 mV/(mW/cm^2)
Amplifier Gains : 20.00 20.00 20.00

Sample:

Rate: 6000 Samples/Sec
Count: 100 Samples
NIDAQ Gain: 5

Comments:

1900 MHz Validation

CF=1; Amb. Temp= 22.4 'C; Liq. Temp=22.1 'C

Area Scan - Max Peak SAR Value at x=2.0 y=-3.0 = 9.10 W/kg

Zoom Scan - Max Peak SAR Value at x=2.0 y=-3.0 z=0.0 = 16.81 W/kg


Max 1g SAR at x=2.0 y=-3.0 z=0.0 = 9.76 W/kg

Max 10g SAR at x=2.0 y=-2.0 z=0.0 = 5.09 W/kg

Validation Results at 0.25 W:

Peak Nominal = 18.0, Error: -6.75 %
1g Nominal = 9.9, Error: -1.68 %
10g Nominal = 5.1, Error: -0.59 %

SAR - Z Axis
at Hotspot x:2.0 y:-3.0



1g SAR Values

