

Certification Exhibit

RF Exposure:

FCC 47 CFR Part 2.1093

Future Cardia, Inc.
FC-MX2

Module FCC ID: 2BUWJ-FCICM100

Prepared for: Future Cardia, Inc.
910 Woodbridge Ct
Safety Harbor, FL 34695

COMMERCIAL-IN-CONFIDENCE

Document Number: NC721005002.6 | Issue: 1



America

**Add value.
Inspire trust.**

SIGNATURE

A handwritten signature in black ink, appearing to read "Yves-Thierry Jean-Charles".

NAME	JOB TITLE	RESPONSIBLE FOR	ISSUE DATE
Yves-Thierry Jean-Charles	Senior Engineer	Authorized Signatory	24 February 2026

Signatures in this approval box have checked this document in line with the requirements of TÜV SÜD America, Inc. document control rules.

FCC Accreditation Designation Number US1148 New Brighton, MN Test Laboratory	Innovation, Science, and Economic Development Canada Accreditation Site Number 4512A New Brighton, MN Test Laboratory
--	---

TÜV SÜD America Inc
141 14th Street NW
New Brighton, MN 55112

Phone: 651-631-2487
www.tuv-sud-america.com

TÜV SÜD

TÜV®



Contents

- 1 Report Summary3**
 - 1.1 Document Modification Record3
 - 1.2 Purpose3
 - 1.3 General Information.....3
 - 1.4 Device under Test (DUT) Description3
 - 1.5 Technical Information3
- 2 Evaluation Details4**
 - 2.1 Specific Requirements4
 - 2.2 RF Exposure Assessment.....5
- 3 Conclusion5**



1 Report Summary

1.1 Document Modification Record

Alterations and Additions of this document will be issued to the holders of each copy in the form of a complete document.

Issue	Description of Change	Date of Issue
1	First Issue	2026-February-24

1.2 Purpose

The purpose of the document is to report the RF exposure evaluation of the FC-MX2 in accordance with FCC 47 CFR Part 2.1093.

1.3 General Information

Applicant: Future Cardia, Inc.
Device Category: Portable
Environment: General Population/Uncontrolled Exposure
Test Specification/Issue/Date: US Code of Federal Regulations (CFR): Title 47, Part 1, Subpart I: Procedures Implementing the National Environmental Policy Act of 1969, 2025.
US Code of Federal Regulations (CFR): Title 47, Part 2, Subpart J: Equipment Authorization Procedures, 2025.

1.4 Device under Test (DUT) Description

The EUT is an Insertable Cardiac Monitor System (ICM). Insertable cardiac monitors (ICMs) are implanted under the skin on the left side of a patient's chest to provide continuous long-term heart rhythm monitoring and identify or exclude arrhythmias as the cause of symptoms. The EUT includes a Bluetooth Low Energy (BLE) radio.

1.5 Technical Information

FCC ID: 2BUWJ-FCICM100
Transmission Protocol: Bluetooth Low Energy (BLE)
Frequency: 2402 - 2480 MHz
Antenna Type: Monopole
Antenna Gain: -9.61 dBi
Maximum RF Output Power: -0.46 dBm
Exposure Conditions: ≤ 5 mm (implantable device)



2 Evaluation Details

2.1 Specific Requirements

FCC 47 CFR Part 1.1307 Actions that may have a significant environmental effect, for which Environmental Assessments (EAs) must be prepared.

1.1307 (b)(3) Determination of exemption.

(i) For single RF sources (i.e., any single fixed RF source, mobile device, or portable device, as defined in paragraph (b)(2) of this section): A single RF source is exempt if:

(A) The available maximum time-averaged power is no more than 1 mW, regardless of separation distance. This exemption may not be used in conjunction with other exemption criteria other than those in paragraph (b)(3)(ii)(A) of this section. Medical implant devices may only use this exemption and that in paragraph (b)(3)(ii)(A);

(B) Or the available maximum time-averaged power or effective radiated power (ERP), whichever is greater, is less than or equal to the threshold P_{th} (mW) described in the following formula. This method shall only be used at separation distances (cm) from 0.5 centimeters to 40 centimeters and at frequencies from 0.3 GHz to 6 GHz (inclusive). P_{th} is given by:

$$P_{th} = \begin{cases} ERP_{20\text{ cm}} \left(\frac{d}{20\text{ cm}} \right)^x & d \leq 20\text{ cm} \\ ERP_{20\text{ cm}} & 20\text{ cm} < d \leq 40\text{ cm} \end{cases}$$

Where

$$x = -\log_{10} \left(\frac{60}{ERP_{20\text{ cm}} \sqrt{f}} \right) \quad \text{and } f \text{ is in GHz};$$

And

$$ERP_{20\text{ cm}} (\text{mw}) = \begin{cases} 2040f & 0.3\text{ GHz} \leq f < 1.5\text{ GHz} \\ 3060 & 1.5\text{ GHz} \leq f < 6\text{ GHz} \end{cases}$$

$d = \text{the separation distance (cm)}$

(C) Or using Table 1 and the minimum separation distance (R in meters) from the body of a nearby person for the frequency (f in MHz) at which the source operates, the ERP (watts) is no more than the calculated value prescribed for that frequency. For the exemption in Table 1 to apply, R must be at least $\lambda/2\pi$, where λ is the free-space operating wavelength in meters. If the ERP of a single RF source is not easily obtained, then the available maximum time-averaged power may be used in lieu of ERP if the physical dimensions of the radiating structure(s)



do not exceed the electrical length of $\lambda/4$ or if the antenna gain is less than that of a half-wave dipole (1.64 linear value).

Table 1 to § 1.1307(b)(3)(i)(C)—Single RF Sources Subject to Routine Environmental Evaluation

RF Source frequency (MHz)	Threshold ERP (watts)
0.3-1.34	1,920 R ² .
1.34-30	3,450 R ² /f ² .
30-300	3.83 R ² .
300-1,500	0.0128 R ² f.
1,500-100,000	19.2R ² .

2.2 RF Exposure Assessment

Minimum Test Separation Distance: 0.5 cm
 Highest Operating Frequency: 2480.0 MHz
 Antenna Type: Monopole
 Antenna Gain: -9.61 dBi
 Maximum RF Conducted Power: -0.46 dBm
 Maximum RF Conducted Power: 0.8995 mW
 Maximum Transmitter EIRP: -10.07 dBm
 Maximum Transmitter EIRP: 0.098401 mW
 Exposure Condition: Body

Determination of Exemption:

47 CFR 1.1307(b)(3)(i)(A)

Exclusion Threshold: 1.0 mW
 Exemption: YES

Maximum Transmitter ERP: 0.0599791 mW

3 Conclusion

Exemption from Routine Environmental Evaluation applies. The EUT meets the RF exposure requirements of the standards listed above.